

Volume Range	Increment	Test Volume ( $\mu\text{l}$ )	Accuracy error		Precision error	
			%	$\mu\text{l}$	%	$\mu\text{l}$
0.1-2.5 $\mu\text{l}$	0.05 $\mu\text{l}$	2.5	2.50%	0.0625	2.00%	0.05
		1.25	3.00%	0.0375	3.00%	0.0375
		0.25	12.00%	0.03	6.00%	0.015
		10	1.00%	0.1	0.80%	0.08
0.5-10 $\mu\text{l}$	0.1 $\mu\text{l}$	5	1.50%	0.075	1.50%	0.075
		1	2.50%	0.025	1.50%	0.015
		20	0.90%	0.18	0.40%	0.08
2-20 $\mu\text{l}$	0.5 $\mu\text{l}$	10	1.20%	0.12	1.00%	0.1
		2	3.00%	0.06	2.00%	0.04
		50	0.60%	0.3	0.30%	0.15
5-50 $\mu\text{l}$	0.5 $\mu\text{l}$	25	0.90%	0.225	0.60%	0.15
		5	2.00%	0.1	2.00%	0.1
		100	0.80%	0.8	0.15%	0.15
10-100 $\mu\text{l}$	1 $\mu\text{l}$	50	1.00%	0.5	0.40%	0.2
		10	3.00%	0.3	1.50%	0.15
20-200 $\mu\text{l}$	1 $\mu\text{l}$	200	0.60%	1.2	0.15%	0.3
		100	0.80%	0.8	0.30%	0.3
		20	3.00%	0.6	1.00%	0.2
50-200 $\mu\text{l}$	1 $\mu\text{l}$	200	0.60%	1.2	0.15%	0.3
		100	0.80%	0.8	0.30%	0.3
		50	1.00%	0.5	0.40%	0.2
100-1000 $\mu\text{l}$	5 $\mu\text{l}$	1000	0.60%	6	0.20%	2
		500	0.70%	3.5	0.25%	1.25
		100	2.00%	2	0.70%	0.7